


तार : फोरकाष्ट बुर्ला

भारत सरकार

दूरभाष 066-2430238



केन्द्रीय जल आयोग

फैक्स- 0663-2431809

कार्यालय ,कार्यपालक अभियंता

महानदी मंडल,डाक्टर कलोनी बुर्ला, (उड़ीसा.)

पत्र सं-एम डी/एच.क्यू-/एस.ओ-02./2011-12/864-69

दिनांक -19.09.2011

निविदा आमंत्रण सूचना

भारत के राष्ट्रपति की ओर से,अधिशासी अभियंता, महानदी मंडल, केन्द्रीय जल आयोग बुर्ला द्वारा, निम्नलिखित सामानों की आपूर्ति के लिये पंजीकृत एवं प्रतिष्ठित फर्मों से निम्न शर्तों के अनुसार दिनांक 29-09-2011 अपराह्न 03-30 तक मुहरबन्द निविदायें आमंत्रित करता है। निविदायें उसी दिन अपराह्न 04-00 बजे निविदाकर्ता अथवा उनके प्रतिनिधियों (यदि कोई हों) के समक्ष खोले जायेंगे।

क्र. सं.	सामानों का विवरण Description of the item	अनुमानित मात्रा
1.	Hot Air Oven (specification as per enclosed annexure).	One

नियम व शर्तें

- 1.मुहरबन्द लिफाफा (जिसमे निविदा भेजा जा रहा है)के उपर के लिये निविदा Hot Air Oven लिखा होना चाहिये तथा अधिशासी अभियंता, महानदी मंडल, डाक्टर कालोनी, बुर्ला 768017-जिला सम्बलपुर (उड़ीसा) के पते पर भेजना चाहिये।
2. दरें अंकों एवं शब्दों दोनों में लिखना चाहिये। काट-छांट किये गये स्थानों पर तारीख सहित आद्यक्षर करके पुनः लिखना चाहिये।
3. निविदा में दिये गये दरें, निविदा खोलने के कम से कम 30 दिनों तक वैध होनी चाहिये।
4. सामानों की सुपुर्दगी अधोहस्ताक्षरकर्ता के कार्यालय में करना है। अतः दरें कार्यालय तक पहुंचाकर होनी चाहिये। ढुलाई (परिवहन)के दौरान टुट-फुट की जिम्मेदारी आपूर्तिकर्ता की होगी।
5. निविदा के साथ अधिकृत विक्रेता प्रमाणपत्र और वेट चुकता प्रमाणपत्र संलग्न करना चाहिये।
6. निविदा में फर्म का व्यापारी सूची संख्या / स्थायी लेखा संख्या लिखना अनिवार्य है।
7. निविदा में प्रस्तावित सामानों का विस्तृत विवरण एवं Make and Model संख्या लिखना अनिवार्य है अन्यथा निविदा रद्द किये जा सकते हैं।
8. सामानों की आपूर्ति आदेश के दिन से 15 दिन के भीतर करना होगा।
9. सामान्यतः सबसे कम दरें वाली निविदा स्वीकार किये जाते हैं, परन्तु अधोहस्ताक्षरकर्ता किसी या सभी निविदाओं को बिना कारण बताए निरस्त करने का अधिकार सुरक्षित रखता है।
10. आपूर्ति के संबंध में उत्पन्न किसी भी तरह के विवाद की स्थिति में, मामले को अधिक्षण अभियंता, जल वैज्ञानिक प्रेक्षण परिमंडल, केन्द्रीय जल आयोग, भुवनेश्वर को भेजा जायेगा और उनका निर्णय अंतिम होगा एवं दोनों पक्ष मानने को बाध्य होंगे।
- 11.बिल का भुगतान सामान आपूर्ति करने अथवा बिल प्रस्तुत करने (जो बाद में हो)के पश्चात 20 दिनों के भीतर किया जायेगा।
12. क्रेता द्वारा उन्हीं निविदाओं पर विचार किया जायेगा जो सभी नियम एवं शर्तों को पूरा करते हों एवं ठीक से हस्ताक्षर किये हों।
- 13.सभी शुल्क, कर (बिक्रीकर एवं चुंगीकर को छोड़कर) एवं अन्य प्रभार जो देय हों उन्हें कुल रकम के साथ मिलाकर निविदा भरनी चाहिए। आवश्यक होने पर चुंगीकर छूट प्रमाणपत्र इस कार्यालय द्वारा दिया जायेगा, बिक्रीकर यदि कोई हो तो उसे निविदा में अलग से दर्शाना है बिक्रीकर को निविदाओं की तुलना के लिये विचार नहीं किया जायेगा।

(आर.पी.शुक्ला)

अधिशासी अभियंता

महानदी मंडल केन्द्रीय जल आयोग, बुर्ला

प्रतिलिपि:

1. अधीक्षण अभियंता, जल वैज्ञानिक प्रेक्षण परिमंडल, केन्द्रीय जल आयोग, भुवनेश्वर।
2. सूचना पटल महानदी मंडल केन्द्रीय जल आयोग, बुर्ला।
3. निदेशक साफ्टवेयर प्रबंधन निदेशालय, केन्द्रीय जल आयोग, 628, सेवा भवन (दक्षिण) आर.के.पुरम नई दिल्ली 110066।
4. सहायक लेखाधिकारी, लेखा शाखा, केन्द्रीय जल आयोग, बुर्ला।
5. सादस्य लोकल पर्चेस कामेटी उप माडलीय अभियंता मध्य महानदी मंडल केन्द्रीय जल आयोग, बुर्ला।
6. सादस्य लोकल पर्चेस कामेटी उप माडलीय अभियंता आपर महानदी मंडल केन्द्रीय जल आयोग, रायपुर।


(आर.पी.शुक्ला)

pdfMachine

A pdf writer that produces quality PDF files with ease!

Produce quality PDF files in seconds and preserve the integrity of your original documents. Compatible across nearly all Windows platforms, simply open the document you want to convert, click "print", select the "Broadgun pdfMachine printer" and that's it! Get yours now!

**GRAMS: FORECAST
BURLA**



Phone: 0663-2430238

Fax: 0663-2431809

**GOVERNMENT OF INDIA
CENTRAL WATER COMMISSION
OFFICE OF THE EXECUTIVE ENGINEER
MAHANADI DIVISION, DOCTOR'S COLONY
BURLA-768017, Dist: SAMBALPUR (ORISSA)**

No: MD/HQ/SO-02/2011-12/864-69

Dated. 19.09.2011

NOTICE INVITING QUOTATION

For and on behalf of the President of India, the Executive Engineer, Mahanadi Division, Central Water Commission, Burla invites the sealed quotations for the supply of following items from the reputed firms as per the terms and conditions mentioned below up to 03.30 P.M. on 29.09.2011 and the quotations will be opened on the same day at 4.00 P.M. in the presence of representatives of the firms if any.

Sr. No.	सामानों का विवरण Description of the item	अनुमानित मात्रा Tentative Quantity
1.	Hot Air Oven (specification as per enclosed annexure-1).	One

TERMS AND CONDITIONS:

1. The sealed envelope containing the quotation should be super-scribed as "Quotation for supply of " **Hot Air Oven** " and addressed to the Executive Engineer, Mahanadi Division, Central Water Commission, Doctor's Colony, Burla-Sambalpur – 7
2. The rate should be indicated in both figures and words. Corrections/ overwriting if any in the quotation shall be made by initialing, dating and re-writing.
3. The rate quoted shall be valid for a period of 30 days from the date of opening of the quotation.
4. The price shall be inclusive of transportation. Any damage occurred during transit should be borne by the agency itself.
5. Authorized dealership certificates and VAT Clearance certificates may be enclosed along with quotations.
6. TIN/PAN number of the firm must be mentioned on the quotation of the firm.
7. Make and Model number of the proposed materials must be mentioned while quoting the rates. Quotation will be liable for rejection if the rates are quoted without mentioning make and model number of the proposed materials.
8. Materials are to be supplied to the undersigned within 15 days from the date of issue of supply order.
9. Normally, the lowest rates quoted will be accepted but the undersigned reserves the right to reject any or all the quotations without assigning any reasons.
10. Any dispute arising in this supply will be referred to Superintending Engineer, Hydrological Observation Circle, Central Water Commission, Bhubaneswar and his decision will be final and binding on both the parties.
11. Payment will be made within 20 days after full supply of materials, whichever is later, through Demand Draft (after deducting DD Commission) payable at near by SBI branch for outstation agencies/ crossed cheque payable at SBI, Burla branch for local agencies.
12. The purchaser will evaluate and compare the quotations determined to be substantially responsive i.e. which are properly signed; and conform to the terms and conditions and specifications.
13. All duties taxes (except sales tax and octroi) and other levies payable by the supplier/contractor under the contract should be included in the total price. Octroi exemption certificate if necessary will be issued by this office. The sales tax if any may be indicated separately in the quotations this will not be considered for comparison of quotations.

(R.P.Shukla)
Executive Engineer

Copy to

1. The Superintending Engineer, Hydrological Observation Circle, CWC, Bhubaneswar.
2. The Notice Board.
3. The Director, Software Management Directorate, 628, Sewa Bhawan (South), R.K.Puram, New Delhi-110066 along with soft copy sent through email at smdte.nic.in for publishing in the CWC website.
4. The Accounts Branch, MD, CWC, Burla.
5. The Member, Local Purchase committee for SDE, UMSD, Raipur.
6. The Member, Local Purchase committee for SDE, MMSD-II, Burla.

(R.P.Shukla)

pdfMachine

A pdf writer that produces quality PDF files with ease!

Produce quality PDF files in seconds and preserve the integrity of your original documents. Compatible across nearly all Windows platforms, simply open the document you want to convert, click "print", select the "Broadgun pdfMachine printer" and that's it! Get yours now!

Annexure-1

Sr. No.	Description and specifications for Hot Air Oven	
1.	<p>Hot Air Oven of size:14”X14”X14” Specification Wattage:1000W Supply: 220V/230V normal domestic environment Body: Double walled, inside non rustable SS304 sheet duly anodized, outside MS Body duly finished with stove enamel. 2.5” gap between the walls is filled with blanket of high efficiency mineral wool. The door shall be fitted with chromium plated brass hinges with a spring and roller type latch. Three heating element made from the best quality nickel chromium wire with very long life for refractory supports. Covered heaters shall be placed under ribs at the bottom and side for uniform distribution of heat. Two pilot lamps and perforated shelves in side the chamber.</p>	